# Tunable dynamical tissue phantom for laser speckle imaging

SOUMYAJIT SARKAR,<sup>1,\*</sup> K. MURALI,<sup>1,\*</sup> AND HARI M. VARMA<sup>1,#</sup>

<sup>1</sup>Dept of Bio-sciences and Bio-engineering, Indian Institute Technology - Bombay, Mumbai, India - 400076 \* Equal contribution

<sup>#</sup>harivarma@iitb.ac.in

## Abstract:

We introduce a novel method to design and implement a tunable dynamical tissue phantom for laser speckle-based *in-vivo* blood flow imaging. This approach relies on Stochastic Differential Equations (SDE) to control a piezoelectric actuator which, upon illuminated with a laser source, generates speckles of pre-defined probability density function and auto-correlation. The validation experiments show that the phantom can generate dynamic speckles that closely replicate both surfaces as well as deep tissue blood flow for a reasonably wide range and accuracy.

# 1. Introduction

Blood flow plays a crucial role in the human body, serving as a vital conduit for delivering oxygen to tissues and facilitating the removal of metabolic waste. It acts as a key bio-marker in medical diagnostics and research. To measure blood flow non-invasively, methods such as Laser Doppler Flowmetry (LDF) [1,2], Laser Speckle Contrast Imaging (LSCI) [3,4] and Diffuse Correlation Spectroscopy (DCS) [5–7] are widely used. LSCI and its variants including multi-exposure speckle imaging (MESI) use uniform illumination in the region of interest and single-exposure or multi-exposure speckle contrast to measure surface *in-vivo* blood flow changes (up to about 1 mm deep). On the other hand, DCS and its variants [8–10], such as Speckle Contrast Optical Spectroscopy (SCOS) [11–13] and multi speckle DCS [14–16] are employed to measure deep tissue blood flow. Additionally, interferometric detection methods are also employed to boost the low light signal at the detector end [17–20].

All these modalities require tissue-mimicking phantoms for optimizing and calibration purposes to ensure consistent and comparable results across various systems and over time. Currently, liquid phantoms [21–23] made using a combination of Intralipid, glycerol, and water are routinely used to replicate blood flow. In addition, to replicate a directed flow, the liquid phantoms are flown using a pump. However, they face challenges like short shelf life, difficulty in replicating arbitrary geometry, and risk of contamination. On the other hand, solid phantoms can last longer, have consistent optical properties, and are easier to handle [24,25]. However, they fail to replicate the dynamic properties of blood flow. In this context, we propose a new phantom to replicate the dynamic properties of tissues, especially blood flow. Other optical properties of the phantom will be ensured by the nature of the solid PDMS-based phantom itself. We validate the proposed phantom by mimicking the blood flow changes by the phantom for both surface and deep tissue blood flow changes associated with MESI and DCS systems.

Recently, we introduced a laser speckle simulation method based on the Stochastic Differential Equation (SDE) [26] to obtain laser speckles with a predetermined Probability Density Function (PDF) and auto-correlation  $(g_1)$ . In the current study, we extend this methodology to develop a tunable dynamic phantom, capable of simulating various tissue blood flow scenarios. The core idea is to use the SDE-derived signals to actuate a piezo-actuator device to which a diffuser is attached. When a laser source illuminates this phantom, it produces dynamic speckles. By adjusting the parameters of the SDE, we can fine-tune the speckle patterns to represent different blood flow conditions. Depending on the laser source (uniform or point source) and the diffuser employed, our phantom can replicate either superficial or deep tissue blood flow.

# 2. Methods

# 2.1. Construction of proposed tunable phantom



Fig. 1. Schematic diagram of the tunable dynamical phantom.

Fig. 1 illustrates the schematic of the tunable dynamical tissue phantom. The signal (x(t)) as shown in Fig. 1d, is generated using a Stochastic Differential Equation (SDE), as outlined in Ref [26], with a predetermined auto-correlation (Fig. 1(a)) and PDF (Fig. 1(b)). Briefly, a general SDE is given by [27]

$$dx(t) = a(x(t), t)dt + b(x(t), t)dW(t)$$
(1)

with initial condition  $x(0) = x_0$ . As given in our earlier work in Ref [26], we denote *a* and *b* as the drift and diffusion coefficients, respectively. *W* represents the Wiener process (or Brownian motion) with its increments defined as  $dW(t) \sim \sqrt{dt}N(0, 1)$ , where *N* is a Gaussian random variable with zero mean and unit variance. The parameters  $a = -\alpha(x - \mu)$  and  $b = \sqrt{2\mu\alpha x}$  are deduced from the Fokker-Planck equation to ensure that x(t) obeys exponential distribution and exponential auto-correlation.  $\alpha$  is the inverse of the characteristic decay time  $\tau_{SDE}$  which can be related to blood flow velocity [4] and  $\mu$  is mean of x(t).

The signal x(t), obtained by solving SDE in Eq.(1), is utilized to drive a piezoelectric actuator. This actuator vibrates in response to the signal, thereby replicating the behavior of dynamic scatterers within the tissue. A diffuser (made using PDMS,  $TiO_2$  and Indian Ink), tailored to replicate the static optical properties of tissue, is positioned on top of the piezoelectric actuator. When a laser illuminates this assembly, it generates dynamic speckles. These speckles are then measured using either LSCI or DCS systems, to characterize the proposed phantom.

# 2.2. Signal generation and Actuation

We solved SDE numerically using the Milstein scheme in Matlab <sup>®</sup> for different  $\tau_{SDE}$  and multiple realizations [26]. These simulated stochastic time series x(t) (shown in Fig. 1c) are then converted to analog voltages using a data acquisition card (DAQ). The signal is then amplified using op-amp based voltage amplifier (Fig. 1e) before feeding to the Piezoelectric actuator (Piezo-Hap 1919, shown in Fig. 1h). A thin (Fig. 1f) or thick diffuser(Fig. 1g) is attached to the piezo-actuator to replicate the superficial or deep tissue blood flow respectively. Thick diffuser is made of Polydimethylsiloxane(PDMS) [21, 28] with optical absorption and reduced scattering coefficients of  $\mu_a = 0.003 cm^{-1}$  and  $\mu'_s = 10 cm^{-1}$  respectively.

In the experiments, we employed two methods, as shown in Fig. 2, to optimize the operation of the actuator. We used a Tektronix AFG1022 generator as a DAQ (Data Acquisition) device



Fig. 2. Timing diagram of Method-1 and Method-2 utilised to operate the actuator.

to convert the signal x(t) into voltage v(t) for driving the actuator. The DAQ device has 255 registers, denoted by Register-n (in Fig. 2), each storing  $N_s = 8000$  data points pre-loaded for independent realizations of x(t). In Method-1, as shown in Fig. 2a,  $N_s$  values are stored within a time interval of  $T_n = 1s$ , resulting in a minimum time interval of dt = 0.125ms. This is followed by a time delay of  $T_{delay} = 0.1s$  before the next register begins operation. In Method-2,  $N_s$  values from a register are stored at regular intervals of  $t_m = 0.1s$  for ten cycles, resulting in  $T_n = 1s$  with a finer time resolution of dt = 0.0125ms. Method-2, as shown in Fig. 2b, enhances the range of flow speeds that can be realized by the phantom. Note that in Method-2, the same values are repeated over multiple cycles.

## 2.3. Experimental Setup - Surface blood flow

We used a diode laser at 785 nm (Holmarc Optomechatronics Pvt Limited, India), to provide uniform illumination over a region of  $2 \times 2$  cm, and employed a CCD-based camera (Basler  $acA640-120\mu m$ ) with a 50 mm objective lens to measure the scattered intensity. Multi-exposure data was captured at 15 exposure times ranging from 100  $\mu s$  to 2512  $\mu s$ , distributed in a log scale. A small region of interest (ROI) of  $1 \times 1$  mm was mapped to a (100 × 100) pixel area of the camera sensor, with appropriate f/# to match the speckle-to-pixel ratio. 100 images were acquired at each exposure time to calculate the multi-exposure speckle contrast data for various  $\tau_{SDE}$  values. The speckle contrast ( $\kappa$ ) was calculated for each pixel over frames and its mean was computed. Three sets of such experiments were performed for individual  $\tau_{SDE}$ values to measure the repeatability. The mean of the trials was fitted to multi-exposure speckle contrast model [4,29] to determine the de-correlation time  $(\tau_{c_k})$ . This process was repeated for multiple flows ( $\tau_{SDE}$ ) to determine the calibration characteristics of the phantom. To show the feasibility of working of the proposed method for *in-vivo* tissue, we have performed a human forearm cuff experiment, wherein the blood flow in human arm was occluded by using a standard sphygmomanometer. This work was undertaken with approval of the Institute Ethical Committee of the Indian Institute of Technology - Bombay, Approval Number: IITB-IEC/2018/017. A single exposure speckle contrast data (measured at 300  $\mu s$  exposure time) was acquired over a period of 180 s over pre-occlusion, occlusion and post-occlusion phases.

# 2.4. Experimental Setup - Deep tissue blood flow

For DCS measurements, we focused the aforementioned laser source to form a 1 mm diameter point source. We employed a single photon avalanche diode (SPAD) detector (SPCM-AQRH-14-FC) in conjunction with a hardware auto-correlator (correlator.com), at a source-detector distance of 2 cm. The normalized intensity auto-correlation  $g_2(\tau)$  was measured for 30 seconds in each trial. The mean values from three independent trials conducted for various  $\tau_{SDE}$  were fitted to the semi-infinite solution to Correlation Diffusion Equation (CDE) [6].

# 3. Results

#### 3.1. Mimicking Surface blood flow



Fig. 3. Replicating superficial blood flow: (a) Normalized multi-exposure Speckle contrast data( $\kappa_n$ ), along with the fit to determine the de-correlation time  $\tau_{C_{\kappa}}$ , for two different flow realizations ( $\tau_{SDE}$ ) obtained from the proposed dynamic phantom. (b) Calibration curve of the phantom showing characteristic decay time ( $\tau_{SDE}$ ) used in the SDE model versus the experimentally obtained decay time ( $\tau_{c_{\kappa}}$ ). The high correlation coefficient between them, along with minimal variance (less than 5% relative change with the mean), was observed over three trials.

In this section, we present the result of dynamic speckles produced to mimic the surface blood flow. The SDE was solved to obtain x(t) for each flow  $(\tau_{SDE})$ , which was used to drive the piezo actuator. Upon illuminating with a uniform laser source, the scattered intensities at multiple exposures were captured for 3 trials and speckle contrast( $\kappa$ ) was calculated as shown in Fig. 3(a). It was along with the fit for analytical solution [4] related to surface blood flow changes. The two representative  $\tau_{SDE}$  of 1.47ms and 0.1995ms, indicated as flow 1 and flow 2 respectively. The fitted decorrelation time ( $\tau_{c_{\kappa}}$ ) for flow 1 and flow 2 were 0.58 ± 0.0175ms and 0.268 ± .0054ms respectively over three trials. In Fig. 3b, the calibration curve of the phantom for the said experiments is shown, demonstrating a linear relationship between  $\tau_{SDE}$  and  $\tau_{c_{\kappa}}$  in the range of 200  $\mu s$  to 2.7 ms, with a high correlation coefficient ( $R^2$ ) of 0.96. Data points indicated in blue were generated using Method-1 while data points indicated in red were generated using Method-2. These two methods were employed to ensure the extended dynamic range of our proposed phantom.

# 3.2. Mimicking Deep tissue blood flow

We present the results of the proposed phantom for replicating deep tissue blood flow using DCS experiments, with a thick diffuser attached to the piezo-actuator. Fig. 4(a) shows the mean  $g_1$  values resulting from three trials conducted with the phantom. The representative measurements shown here were taken for two distinct values of  $\tau_{SDE}$ : 0.58 *ms* and 5 *ms*, denoted as flow 1 and flow 2 respectively. The  $g_1$  was fitted to the semi-infinite solution of CDE, with a reasonable agreement. Furthermore, a calibration curve, akin to the one used in MESI (as shown in Fig.



Fig. 4. Replicating deep tissue blood flow: (a) Mean  $g_1$  values obtained from three experimental trials for two representative flows, along with their fits to semi-infinite solution of CDE. (b) Calibration curve showing a linear relationship between  $D_B$  and  $\tau_{SDE}$  is presented, demonstrating the potential of phantoms for deep tissue blood flow studies.

3(b) is presented in Fig. 4(b). The correlation coefficient ( $R^2$ ) for this curve is 0.92, indicating a strong linear relationship, albeit slightly lower than the MESI curve.



3.3. Mimicking in-vivo human handcuff experiment

Fig. 5. Comparison of single-exposure speckle contrast measurements between the dynamic phantom and *in-vivo* human hand cuff experiments, demonstrating the phantom's wide dynamic range and rapid tunability. The dotted lines indicate the start and end of the blood cuff in the sphygmomanometer.

Fig. 5 shows the comparison of a single exposure handcuff experiment over a phase of 180 sec with three stages of baseline occlusion and post-releasing phase along with the mimicking values realized by the proposed phantom. This depicts the phantoms capability to generate such a wide range of changes in a fast and reliable manner.



Fig. 6. Frequency response analysis of the phantom. (a) The relationship between input frequencies and speckle contrast at multiple exposure times, (b) a zoomed-in version revealing saturation above approximately 6 kHz, indicating the phantom's maximum operating frequency, and (c) Simulation of power spectral density.

To determine the maximum feasible operating frequency of the phantom, we performed a frequency response analysis. For this, we applied varying sinusoidal frequencies as input signals to the proposed phantom and measured the corresponding multi-exposure speckle contrast. Fig. 6(a) illustrates the outcome of this analysis, demonstrating the relationship between input frequencies and speckle contrast. A zoomed-in version of the same is shown in Fig. 6(b), wherein we observed that signals approximately above 6 kHz showed signs of saturation. As a result, we concluded that the piezo-actuator along with the amplifier system we designed, has a maximum operating frequency of 6 kHz. We also conducted a power spectral density analysis using the experimentally acquired  $\tau_{c_k}$  values as shown in Fig. 6(c). We simulated  $g_1$  based on the experimental  $\tau_{c_k}$  values. Subsequently, we calculated its power spectrum to determine the cutoff point, denoted as  $\omega_c$ , which corresponds to the -20dB point from the peak amplitude. The  $\omega_c$  plotted against  $\tau_{c_k}$  as in Fig. 6c, confirms that the phantom's response effectively extends up to 6 kHz.

## 4. Discussion

In this paper, we have designed and developed a tunable solid phantom for laser speckle imaging to measure blood flow. To accomplish this, we employed our recently introduced SDE based algorithm [26] to generate signals x(t) that control the piezo actuator. By leveraging this approach, we simulated a wide range of blood flow changes for both surface and deep tissue speckle-based imaging methods and compared them with *in-vivo* human experiments.

For surface blood flow, we validated our approach by obtaining the multi-exposure speckle contrast (Fig. 3a) and fitting it to an analytical solution. We also plotted the characteristic decay time obtained by the proposed phantom mimicking multiple flows (Fig. 3b), which are in

reasonable agreement with the desired flows, as given for simulation using SDE, with a high correlation coefficient ( $R^2$ ) of 0.96. For the deep tissue blood flow, we obtained the normalized field autocorrelation at a source-detector separation of 2 cm (Fig 4a) and also show the inverse relationship between fitted blood flow ( $D_B$ ) and  $\tau_{SDE}$  (Fig 4b). The correlation coefficient ( $R^2$ ) for the deep tissue flow is 0.92, indicating a strong linear relationship, although it is slightly lower when compared to the surface flow case (Fig 3b). This difference in correlation could be attributed to the use of a simple exponential model in SDE to generate x(t). In our analysis, we acknowledge the potential for improvement by employing a complex exponential model, as proposed in Ref. [26], for generating x(t). We believe that this enhancement in modeling could lead to an increase in the correlation coefficient in deep tissue cases. Nevertheless, these results act as compelling proof of concept, demonstrating the feasibility of using dynamic phantoms to replicate flow conditions in DCS.

To demonstrate the dynamic tunability of our proposed phantom over a broad range, we obtained the speckle contrast during a human handcuff experiment (Fig. 5a). We then used the calibration curve (Fig. 3b) to generate SDE signals to the piezo actuator and obtained similar speckle contrast based on the proposed phantom, which is in reasonable comparison with one another. This indicates that we can fine-tune a large dynamic range of signals based on the proposed method, limited only by the dynamic range of the piezo transducer and frame rate of the DAQ. Further based on the frequency response analysis (Fig 6) of the transducer, we analyzed that the maximum operable frequency of the piezo actuator is 6 kHz, and further this range can be improved by using better piezo-actuators of higher bandwidth.

We note that the primary objective of these tunable tissue phantoms is to replicate the dynamic characteristics of the tissue, such as tissue blood flow modeled as Brownian motion, rather than mimicking optical properties like absorption and scattering coefficients. Presently, these optical properties are reproduced through the implementation of solid diffusers, attached to the piezo-actuator. Secondly, we have not explored different geometries or heterogeneous blood flow patterns. Nonetheless, we would like to note that employing a piezo-actuator array, similar to the ones demonstrated in Ref. [30, 31], holds promise for heterogeneous flow patterns by exerting independent control over each piezo actuator, which could act as individual pixels.

We emphasize the necessity of employing SDE in our work by noting that any random fluctuations alone cannot replicate blood flow without appropriately in-corporating the desired auto-correlation via SDE. For instance, when disregarding auto-correlation by setting a = 0 and b = 1 in the aforementioned SDE equation (1), resulting in x(t) = dW(t) which is Brownian motion alone, the auto-correlation between time-instances, t and  $t + \tau$  yields  $\sqrt{1/(1 + \tau/t)}$ . This relationship, as given in Chapter 4, equation 4.13 in Ref [32], fails to exhibit the exponentially decaying auto-correlation characteristics of blood flow dynamics.

Furthermore, it's important to acknowledge that while our study focused on employing the SDE-based algorithm for signal generation to drive the piezo actuator, alternative methods based on random signals could be explored. Approaches such as those based on COPULA [33, 34], Fourier transform of the phase matrix [35], correlation matrices [36], and Karhunen–Loève expansion of electric field [37] offer promising avenues for further investigation. In comparison to the above methods, our proposed method offers a possibility to incorporate different types of auto-correlation, not restricted to simple exponential auto-correlation. Besides this, one can also incorporate different PDF by deploying Fokker-plank equations as explained in detail in our earlier work [26]. In future, the proposed phantom could be used for comparison across multiple laser speckle imaging modalities [3,5,38–40] and potentially optimizing various parameters such as exposure time and the number of frames/speckles, which are necessary for enhancing both surface and deep tissue blood flow measurement systems.

**Funding.** Wadhwani Research Centre for Bioengineering (WRCB), Indian Institute of Technology Bombay; CRG–Core research grant 2023, Science Engineering Research Board - Department of Science

and Technology, Ministry of Science and Technology, India.

#### Disclosures.

The authors declare no conflicts of interest. Portions of this work were presented at the European Conferences on Biomedical Optics (ECBO), 2023, in Ref [41].

**Data availability.** Data underlying the results presented in this paper are not publicly available at this time but may be obtained from the authors upon reasonable request.

#### References

- V. Rajan, B. Varghese, T. G. van Leeuwen, and W. Steenbergen, "Review of methodological developments in laser doppler flowmetry," Lasers medical science 24, 269–283 (2009).
- T. Binzoni, A. Humeau-Heurtier, P. Abraham, and G. Mahé, "Blood perfusion values of laser speckle contrast imaging and laser doppler flowmetry: is a direct comparison possible?" IEEE Trans. on Biomed. Eng. 60, 1259–1265 (2012).
- J. Briers and A. Fercher, "Retinal blood-flow visualization by means of laser speckle photography." Investig. ophthalmology & visual science 22, 255–259 (1982).
- A. B. Parthasarathy, W. J. Tom, A. Gopal, *et al.*, "Robust flow measurement with multi-exposure speckle imaging," Opt. express 16, 1975–1989 (2008).
- D. A. Boas, L. Campbell, and A. G. Yodh, "Scattering and imaging with diffusing temporal field correlations," Phys. review letters 75, 1855 (1995).
- T. Durduran, R. Choe, W. Baker, and A. G. Yodh, "Diffuse optics for tissue monitoring and tomography," Reports on Prog. Phys. 73, 076701 (2010).
- L. Dong, L. He, Y. Lin, *et al.*, "Simultaneously extracting multiple parameters via fitting one single autocorrelation function curve in diffuse correlation spectroscopy," IEEE Trans. on Biomed. Eng. 60, 361–368 (2012).
- A. Sdobnov, G. Piavchenko, A. Bykov, and I. Meglinski, "Advances in dynamic light scattering imaging of blood flow," Laser & Photonics Rev. p. 2300494 (2023).
- E. James and P. R. Munro, "Diffuse correlation spectroscopy: A review of recent advances in parallelisation and depth discrimination techniques," Sensors 23, 9338 (2023).
- M. Seong, Y. Oh, K. Lee, and J. G. Kim, "Blood flow estimation via numerical integration of temporal autocorrelation function in diffuse correlation spectroscopy," Comput. Methods Programs Biomed. 222, 106933 (2022).
- C. P. Valdes, H. M. Varma, A. K. Kristoffersen, et al., "Speckle contrast optical spectroscopy, a non-invasive, diffuse optical method for measuring microvascular blood flow in tissue," Biomed. optics express 5, 2769–2784 (2014).
- H. M. Varma, C. P. Valdes, A. K. Kristoffersen, *et al.*, "Speckle contrast optical tomography: A new method for deep tissue three-dimensional tomography of blood flow," Biomed. optics express 5, 1275–1289 (2014).
- B. Kim, S. Zilpelwar, E. J. Sie, *et al.*, "Measuring human cerebral blood flow and brain function with fiber-based speckle contrast optical spectroscopy system," Commun. Biol. 6, 844 (2023).
- K. Murali and H. M. Varma, "Multi-speckle diffuse correlation spectroscopy to measure cerebral blood flow," Biomed. Opt. Express 11, 6699–6709 (2020).
- M. A. Wayne, E. J. Sie, A. C. Ulku, *et al.*, "Massively parallel, real-time multispeckle diffuse correlation spectroscopy using a 500× 500 spad camera," Biomed. Opt. Express 14, 703–713 (2023).
- K. Murali, A. Nandakumaran, and H. M. Varma, "On the equivalence of speckle contrast-based and diffuse correlation spectroscopy methods in measuring in vivo blood flow," Opt. Lett. 45, 3993–3996 (2020).
- W. Zhou, M. Zhao, O. Kholiqov, and V. J. Srinivasan, "Multi-exposure interferometric diffusing wave spectroscopy," Opt. letters 46, 4498–4501 (2021).
- J. Xu, A. K. Jahromi, J. Brake, et al., "Interferometric speckle visibility spectroscopy (isvs) for human cerebral blood flow monitoring," APL Photonics 5 (2020).
- Y. X. Huang, S. Mahler, J. Mertz, and C. Yang, "Interferometric speckle visibility spectroscopy (isvs) for measuring decorrelation time and dynamics of moving samples with enhanced signal-to-noise ratio and relaxed reference requirements," Opt. Express 31, 31253–31266 (2023).
- M. Zhao, W. Zhou, S. Aparanji, *et al.*, "Interferometric diffusing wave spectroscopy imaging with an electronically variable time-of-flight filter," Optica 10, 42–52 (2023).
- 21. R. Choe, "Diffuse optical tomography and spectroscopy of breast cancer and fetal brain," Univ. Pa. (2005).
- C. K. McGarry, L. J. Grattan, A. M. Ivory, *et al.*, "Tissue mimicking materials for imaging and therapy phantoms: a review," Phys. Med. & Biol. 65, 23TR01 (2020).
- L. Cortese, G. L. Presti, M. Pagliazzi, *et al.*, "Liquid phantoms for near-infrared and diffuse correlation spectroscopies with tunable optical and dynamic properties," Biomed. optics express 9, 2068–2080 (2018).
- 24. S. K. V. Sekar, A. Pacheco, P. Martella, *et al.*, "Solid phantom recipe for diffuse optics in biophotonics applications: a step towards anatomically correct 3d tissue phantoms," Biomed. optics express **10**, 2090–2100 (2019).
- A. M. Goldfain, P. Lemaillet, D. W. Allen, *et al.*, "Polydimethylsiloxane tissue-mimicking phantoms with tunable optical properties," J. Biomed. Opt. 27, 074706–074706 (2022).
- K. Murali and H. M. Varma, "Laser speckle simulation tool based on stochastic differential equations for bio imaging applications," Biomed. Opt. Express 13, 6745–6762 (2022).

- C. A. Braumann, Introduction to stochastic differential equations with applications to modelling in biology and finance (John Wiley & Sons, 2019).
- R. Paul, K. Murali, S. Chetia, and H. M. Varma, "A simple algorithm for diffuse optical tomography without jacobian inversion," Biomed. Phys. & Eng. Express 8, 045001 (2022).
- K. Murali, A. Nandakumaran, T. Durduran, and H. M. Varma, "Recovery of the diffuse correlation spectroscopy data-type from speckle contrast measurements: towards low-cost, deep-tissue blood flow measurements," Biomed. optics express 10, 5395–5413 (2019).
- L. Schinaia, A. Scorza, F. Botta, et al., "A novel, low-cost and reliable workbench for optimal voltage distribution on piezoelectric array actuators," in Actuators, vol. 8 (MDPI, 2019), p. 6.
- Y. Qiu, J. V. Gigliotti, M. Wallace, *et al.*, "Piezoelectric micromachined ultrasound transducer (pmut) arrays for integrated sensing, actuation and imaging," Sensors 15, 8020–8041 (2015).
- 32. K. Jacobs, Stochastic processes for physicists: understanding noisy systems (Cambridge University Press, 2010).
- D. D. Duncan and S. J. Kirkpatrick, "The copula: a tool for simulating speckle dynamics," JOSA A 25, 231–237 (2008).
- 34. L. K. Frisk, M. Verma, F. Bešlija, et al., "Comprehensive workflow and its validation for simulating diffuse speckle statistics for optical blood flow measurements," Biomed. Opt. Express 15, 875–899 (2024).
- S. J. Kirkpatrick, D. D. Duncan, and E. M. Wells-Gray, "Detrimental effects of speckle-pixel size matching in laser speckle contrast imaging," Opt. letters 33, 2886–2888 (2008).
- L. Song, Z. Zhou, X. Wang, *et al.*, "Simulation of speckle patterns with pre-defined correlation distributions," Biomed. optics express 7, 798–809 (2016).
- 37. E. James, S. Powell, and P. Munro, "Simulation of statistically accurate time-integrated dynamic speckle patterns in biomedical optics," Opt. letters 46, 4390–4393 (2021).
- Q. Fang, A. Tomar, and A. K. Dunn, "Wide-field intensity fluctuation imaging," Biomed. Opt. Express 15, 1004–1020 (2024).
- 39. L. Kreiss, M. Wu, M. Wayne, et al., "Beneath the surface: Revealing deep-tissue blood flow in human subjects with massively parallelized diffuse correlation spectroscopy," arXiv preprint arXiv:2403.03968 (2024).
- M. Seong, "Comparison of numerical-integration-based methods for blood flow estimation in diffuse correlation spectroscopy," Comput. Methods Programs Biomed. 241, 107766 (2023).
- 41. S. Sarkar, K. Murali, and H. M. Varma, "Development of a tunable dynamical tissue phantom for laser speckle imaging of blood flow," in *Diffuse Optical Spectroscopy and Imaging IX*, vol. 12628 (SPIE, 2023), pp. 105–107.